

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121322\
 Data File : VY011785.D
 Acq On : 13 Dec 2022 12:02
 Operator : KP/MD
 Sample : N6022-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-01-121222

Quant Time: Dec 13 22:17:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 04:19:59 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	50118	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	81002	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	70317	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	28849	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	20692	61.617	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 123.240%	
35) Dibromofluoromethane	7.716	113	20022	56.109	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 112.220%	
50) Toluene-d8	10.179	98	51255	42.435	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 84.860%	
62) 4-Bromofluorobenzene	12.483	95	18764	38.945	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 77.880%	
Target Compounds						Qvalue
20) Methylene Chloride	4.729	84	6670	8.858	ug/l	86
80) 1,3,5-Trimethylbenzene	12.812	105	1632	1.033	ug/l	96
84) 1,2,4-Trimethylbenzene	13.117	105	1938	1.246	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121322\
Data File : VY011785.D
Acq On : 13 Dec 2022 12:02
Operator : KP/MD
Sample : N6022-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EO-01-121222

Quant Time: Dec 13 22:17:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 10 04:19:59 2022
Response via : Initial Calibration

