

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121322\
 Data File : VY011787.D
 Acq On : 13 Dec 2022 12:47
 Operator : KP/MD
 Sample : N5964-01
 Misc : 2.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SU-518-COMP-01

Quant Time: Dec 13 22:17:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 04:19:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	87664	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	138586	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	124789	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	54843	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	32419	54.638	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.280%	
35) Dibromofluoromethane	7.716	113	31998	52.134	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 104.260%	
50) Toluene-d8	10.179	98	89073	43.224	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 86.440%	
62) 4-Bromofluorobenzene	12.483	95	34188	41.800	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 83.600%	
Target Compounds						
16) Acetone	3.948	43	6476	27.054	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121322\
Data File : VY011787.D
Acq On : 13 Dec 2022 12:47
Operator : KP/MD
Sample : N5964-01
Misc : 2.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SU-518-COMP-01

Quant Time: Dec 13 22:17:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 10 04:19:59 2022
Response via : Initial Calibration

