

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121322\
 Data File : VY011792.D
 Acq On : 13 Dec 2022 14:40
 Operator : KP/MD
 Sample : N6025-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 NB-07-121222

Quant Time: Dec 13 22:18:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 04:19:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	279484	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	463809	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	392946	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	154756	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	94301	49.387	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.780%
35) Dibromofluoromethane	7.716	113	98551	47.641	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.280%
50) Toluene-d8	10.179	98	282906	40.630	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	81.260%
62) 4-Bromofluorobenzene	12.483	95	101314	36.438	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	72.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121322\
Data File : VY011792.D
Acq On : 13 Dec 2022 14:40
Operator : KP/MD
Sample : N6025-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
NB-07-121222

Quant Time: Dec 13 22:18:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 10 04:19:59 2022
Response via : Initial Calibration

