

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121322\
 Data File : VY011799.D
 Acq On : 13 Dec 2022 17:19
 Operator : KP/MD
 Sample : N6048-01
 Misc : 5.45g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH-22

Quant Time: Dec 13 22:19:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 04:19:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	288300	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	480526	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	420566	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	194661	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	96012	48.677	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.360%
35) Dibromofluoromethane	7.716	113	103322	48.260	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.520%
50) Toluene-d8	10.179	98	307798	43.051	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	86.100%
62) 4-Bromofluorobenzene	12.483	95	119551	42.199	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	84.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121322\
Data File : VY011799.D
Acq On : 13 Dec 2022 17:19
Operator : KP/MD
Sample : N6048-01
Misc : 5.45g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MH-22

Quant Time: Dec 13 22:19:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 10 04:19:59 2022
Response via : Initial Calibration

