

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121322\
 Data File : VY011802.D
 Acq On : 13 Dec 2022 18:27
 Operator : KP/MD
 Sample : N6037-02
 Misc : 6.67g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RB-17-(1-3)

Quant Time: Dec 13 22:19:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 04:19:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	296018	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	491050	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	424865	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	191905	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	98631	48.703	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.400%
35) Dibromofluoromethane	7.715	113	104500	47.721	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.440%
50) Toluene-d8	10.178	98	304362	41.410	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	82.820%
62) 4-Bromofluorobenzene	12.483	95	120852	41.690	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	83.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121322\
Data File : VY011802.D
Acq On : 13 Dec 2022 18:27
Operator : KP/MD
Sample : N6037-02
Misc : 6.67g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RB-17-(1-3)

Quant Time: Dec 13 22:19:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 10 04:19:59 2022
Response via : Initial Calibration

