

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121322\
 Data File : VY011821.D
 Acq On : 14 Dec 2022 03:08
 Operator : KP/MD
 Sample : N6037-15
 Misc : 5.37g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RB-24-(7-9)

Quant Time: Dec 14 05:38:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 04:19:59 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	240479	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	409425	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	352254	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	157760	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	85758	52.500	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 105.000%	
35) Dibromofluoromethane	7.716	113	87085	47.695	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 95.380%	
50) Toluene-d8	10.179	98	247703	40.237	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 80.480%	
62) 4-Bromofluorobenzene	12.483	95	96198	39.573	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 79.140%	
Target Compounds						Qvalue
20) Methylene Chloride	4.722	84	28463	7.141	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121322\
Data File : VY011821.D
Acq On : 14 Dec 2022 03:08
Operator : KP/MD
Sample : N6037-15
Misc : 5.37g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RB-24-(7-9)

Quant Time: Dec 14 05:38:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 10 04:19:59 2022
Response via : Initial Calibration

