

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121322\
 Data File : VY011826.D
 Acq On : 14 Dec 2022 05:01
 Operator : KP/MD
 Sample : N6037-22
 Misc : 4.51g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RB-27-(5-7)

Quant Time: Dec 14 05:38:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 04:19:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	251000	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	433386	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	361699	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	144276	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	92905	54.692	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.380%	
35) Dibromofluoromethane	7.722	113	97337	50.598	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 101.200%	
50) Toluene-d8	10.179	98	275386	42.646	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 85.300%	
62) 4-Bromofluorobenzene	12.483	95	96351	37.174	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 74.340%	
Target Compounds						
20) Methylene Chloride	4.723	84	30633	7.571	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121322\
Data File : VY011826.D
Acq On : 14 Dec 2022 05:01
Operator : KP/MD
Sample : N6037-22
Misc : 4.51g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RB-27-(5-7)

Quant Time: Dec 14 05:38:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 10 04:19:59 2022
Response via : Initial Calibration

