

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121422\
 Data File : VY011837.D
 Acq On : 14 Dec 2022 12:10
 Operator : KP/MD
 Sample : N6046-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-01-121322

Quant Time: Dec 15 01:18:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 04:19:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	3533	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	5206	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	3867	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	685	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	1255	52.274	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 104.540%	
35) Dibromofluoromethane	7.710	113	668	27.099	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 54.200%	
50) Toluene-d8	10.185	98	3306	42.615	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 85.220%	
62) 4-Bromofluorobenzene	12.477	95	833	25.346	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 50.700%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121422\
Data File : VY011837.D
Acq On : 14 Dec 2022 12:10
Operator : KP/MD
Sample : N6046-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HR-01-121322

Quant Time: Dec 15 01:18:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 10 04:19:59 2022
Response via : Initial Calibration

