

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121422\
 Data File : VY011845.D
 Acq On : 14 Dec 2022 15:11
 Operator : KP/MD
 Sample : N6037-16
 Misc : 5.52g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RB-25-(1-3)

Quant Time: Dec 15 01:20:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 04:19:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	295768	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	492270	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	397477	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	141452	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	114528	57.461	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 114.920%	
35) Dibromofluoromethane	7.716	113	113975	52.290	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 104.580%	
50) Toluene-d8	10.179	98	355948	49.584	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 99.160%	
62) 4-Bromofluorobenzene	12.483	95	111387	37.924	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 75.840%	
Target Compounds						
					Qvalue	
18) Methyl Acetate	4.479	43	21540	9.785	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121422\
Data File : VY011845.D
Acq On : 14 Dec 2022 15:11
Operator : KP/MD
Sample : N6037-16
Misc : 5.52g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RB-25-(1-3)

Quant Time: Dec 15 01:20:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 10 04:19:59 2022
Response via : Initial Calibration

