

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121422\
 Data File : VY011846.D
 Acq On : 14 Dec 2022 15:34
 Operator : KP/MD
 Sample : N6037-20RE
 Misc : 5.82g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RB-26-(2-4)RE

Quant Time: Dec 15 01:20:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 04:19:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	47531	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	74612	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	63490	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	23553	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	18387	57.399	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.800%
35) Dibromofluoromethane	7.722	113	17775	53.926	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.860%
50) Toluene-d8	10.179	98	46160	41.319	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	82.640%
62) 4-Bromofluorobenzene	12.483	95	15739	35.015	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	70.020%
Target Compounds						
					Qvalue	
18) Methyl Acetate	4.473	43	1326	2.623	ug/l #	90
20) Methylene Chloride	4.723	84	5389	6.560	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121422\
Data File : VY011846.D
Acq On : 14 Dec 2022 15:34
Operator : KP/MD
Sample : N6037-20RE
Misc : 5.82g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RB-26-(2-4)RE

Quant Time: Dec 15 01:20:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 10 04:19:59 2022
Response via : Initial Calibration

