

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121422\
 Data File : VY011849.D
 Acq On : 14 Dec 2022 16:42
 Operator : KP/MD
 Sample : N6038-02
 Misc : 1.92g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RB-29-(8-10)

Quant Time: Dec 15 01:21:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 04:19:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	189428	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	315893	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	284017	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	133490	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	68891	53.645	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 107.280%	
35) Dibromofluoromethane	7.715	113	69617	49.569	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 99.140%	
50) Toluene-d8	10.178	98	194805	41.162	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 82.320%	
62) 4-Bromofluorobenzene	12.483	95	77471	41.526	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 83.060%	
Target Compounds						
18) Methyl Acetate	4.478	43	4419	1.895	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121422\
Data File : VY011849.D
Acq On : 14 Dec 2022 16:42
Operator : KP/MD
Sample : N6038-02
Misc : 1.92g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RB-29-(8-10)

Quant Time: Dec 15 01:21:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 10 04:19:59 2022
Response via : Initial Calibration

