

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121422\
 Data File : VY011855.D
 Acq On : 14 Dec 2022 18:57
 Operator : KP/MD
 Sample : N6038-08
 Misc : 3.67g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RB-32-(7-9)

Quant Time: Dec 15 01:23:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 04:19:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	253735	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	429820	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	379729	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	174197	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	87435	50.551	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.100%
35) Dibromofluoromethane	7.716	113	89864	46.809	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.620%
50) Toluene-d8	10.179	98	256982	39.674	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	79.340%
62) 4-Bromofluorobenzene	12.483	95	101263	39.694	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	79.380%
Target Compounds						
					Qvalue	
18) Methyl Acetate	4.479	43	15250	7.757	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121422\
Data File : VY011855.D
Acq On : 14 Dec 2022 18:57
Operator : KP/MD
Sample : N6038-08
Misc : 3.67g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RB-32-(7-9)

Quant Time: Dec 15 01:23:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 10 04:19:59 2022
Response via : Initial Calibration

