

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121422\
 Data File : VY011856.D
 Acq On : 14 Dec 2022 19:20
 Operator : KP/MD
 Sample : N6038-09
 Misc : 6.26g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RB-33-(2-4)

Quant Time: Dec 15 01:23:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 10 04:19:59 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	251435	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	438739	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	382041	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	167031	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	88971	52.052	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 104.100%	
35) Dibromofluoromethane	7.716	113	93570	47.834	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 95.660%	
50) Toluene-d8	10.179	98	265531	40.254	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 80.500%	
62) 4-Bromofluorobenzene	12.483	95	101637	38.947	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 77.900%	
Target Compounds						Qvalue
20) Methylene Chloride	4.716	84	33499	8.875	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121422\
Data File : VY011856.D
Acq On : 14 Dec 2022 19:20
Operator : KP/MD
Sample : N6038-09
Misc : 6.26g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RB-33-(2-4)

Quant Time: Dec 15 01:23:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 10 04:19:59 2022
Response via : Initial Calibration

