


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121422\
 Data File : VY011832.D
 Acq On : 14 Dec 2022 09:25
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y120922S.M
 Title : SW846 8260
 Last Update : Sat Dec 10 04:19:59 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.4	15246	PASS
75	95	30	60	52.3	43413	PASS
95	95	100	100	100.0	83013	PASS
96	95	5	9	7.0	5775	PASS
173	174	0.00	2	1.1	799	PASS
174	95	50	100	86.3	71608	PASS
175	174	5	9	7.1	5096	PASS
176	174	95	101	95.0	68053	PASS
177	176	5	9	7.2	4903	PASS