

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121525\
 Data File : VY023964.D
 Acq On : 15 Dec 2025 09:18
 Operator : SY/MD
 Sample : VY1215SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1215SBL01

Quant Time: Dec 16 03:46:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121025S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 11 03:57:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.726	168	905697	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.628	114	1468696	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.426	117	1218369	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.359	152	499041	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.079	65	382474	47.763	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.520%
35) Dibromofluoromethane	7.652	113	449704	46.841	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.680%
50) Toluene-d8	10.122	98	1624779	45.455	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	90.920%
62) 4-Bromofluorobenzene	12.414	95	475434	39.828	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	79.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121525\
Data File : VY023964.D
Acq On : 15 Dec 2025 09:18
Operator : SY/MD
Sample : VY1215SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1215SBL01

Quant Time: Dec 16 03:46:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121025S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 11 03:57:51 2025
Response via : Initial Calibration

