

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121525\
 Data File : VY023973.D
 Acq On : 15 Dec 2025 13:50
 Operator : SY/MD
 Sample : Q3859-01RE
 Misc : 6.63g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HD-02-121225RE

Quant Time: Dec 16 02:15:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121025S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 11 03:57:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	45310	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.628	114	58846	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.426	117	43863	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	9016	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	18526	46.245	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	92.480%
35) Dibromofluoromethane	7.646	113	20051	52.125	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.260%
50) Toluene-d8	10.115	98	63021	44.004	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	88.000%
62) 4-Bromofluorobenzene	12.414	95	12294	25.705	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	51.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121525\
Data File : VY023973.D
Acq On : 15 Dec 2025 13:50
Operator : SY/MD
Sample : Q3859-01RE
Misc : 6.63g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HD-02-121225RE

Quant Time: Dec 16 02:15:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121025S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 11 03:57:51 2025
Response via : Initial Calibration

