


Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv121619\
 Data File : VY000944.D
 Acq On : 16 Dec 2019 09:34
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y112719S.M
 Title : SW846 8260
 Last Update : Wed Nov 27 12:40:47 2019



AutoFind: Scans 1761, 1762, 1763; Background Corrected with Scan 1752

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.8	62898	PASS
75	95	30	60	49.2	184192	PASS
95	95	100	100	100.0	374122	PASS
96	95	5	9	6.6	24544	PASS
173	174	0.00	2	0.8	2367	PASS
174	95	50	100	76.8	287381	PASS
175	174	5	9	6.7	19309	PASS
176	174	95	101	97.4	279914	PASS
177	176	5	9	6.9	19389	PASS