

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121623\
 Data File : VY016669.D
 Acq On : 16 Dec 2023 15:16
 Operator : SY/MD
 Sample : 05611-07RE
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PE-7RE

Quant Time: Dec 18 00:40:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121523S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 16 00:39:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	6572	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	10684	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	7782	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	1769	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	3138	58.362	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.720%
35) Dibromofluoromethane	7.710	113	3025	52.785	ug/l	-0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.560%
50) Toluene-d8	10.191	98	11731	51.839	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.680%
62) 4-Bromofluorobenzene	12.483	95	2354	32.053	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	64.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121623\
Data File : VY016669.D
Acq On : 16 Dec 2023 15:16
Operator : SY/MD
Sample : 05611-07RE
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PE-7RE

Quant Time: Dec 18 00:40:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121523S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 16 00:39:52 2023
Response via : Initial Calibration

