

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121623\
 Data File : VY016672.D
 Acq On : 16 Dec 2023 16:26
 Operator : SY/MD
 Sample : 05758-37
 Misc : 4.12g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SP-COMP-5

Quant Time: Dec 18 00:41:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121523S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 16 00:39:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	155530	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	266334	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	329327	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	134153	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	82939	65.141	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	130.280%	
35) Dibromofluoromethane	7.728	113	95086	66.559	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	133.120%	
50) Toluene-d8	10.185	98	300421	53.199	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	106.400%	
62) 4-Bromofluorobenzene	12.483	95	92722	50.646	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	101.300%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.729	84	18693	7.224	ug/l	# 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121623\
Data File : VY016672.D
Acq On : 16 Dec 2023 16:26
Operator : SY/MD
Sample : 05758-37
Misc : 4.12g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SP-COMP-5

Quant Time: Dec 18 00:41:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121523S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 16 00:39:52 2023
Response via : Initial Calibration

