

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121623\
 Data File : VY016674.D
 Acq On : 16 Dec 2023 17:13
 Operator : SY/MD
 Sample : 05758-01RE
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SP-COMP-1

Quant Time: Dec 18 00:42:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121523S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 16 00:39:52 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	103915	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	215971	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	279593	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	113980	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	63406	74.487	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 148.980%	
35) Dibromofluoromethane	7.728	113	66405	57.322	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 114.640%	
50) Toluene-d8	10.185	98	336367	72.665	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 145.340%#	
62) 4-Bromofluorobenzene	12.483	95	106037	71.425	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 142.860%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.729	84	16046	10.957	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121623\
Data File : VY016674.D
Acq On : 16 Dec 2023 17:13
Operator : SY/MD
Sample : 05758-01RE
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SP-COMP-1

Quant Time: Dec 18 00:42:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121523S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 16 00:39:52 2023
Response via : Initial Calibration

