

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121623\
 Data File : VY016678.D
 Acq On : 16 Dec 2023 18:47
 Operator : SY/MD
 Sample : 05765-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RT-SB-01-10-11

Quant Time: Dec 18 00:43:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121523S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 16 00:39:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	35574	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	61156	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	55843	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	26095	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	18098	62.161	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.320%
35) Dibromofluoromethane	7.728	113	20904	63.725	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	127.440%
50) Toluene-d8	10.185	98	71637	55.166	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	110.340%
62) 4-Bromofluorobenzene	12.489	95	18427	43.833	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	87.660%
Target Compounds						Qvalue
20) Methylene Chloride	4.729	84	6846	15.104	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121623\
Data File : VY016678.D
Acq On : 16 Dec 2023 18:47
Operator : SY/MD
Sample : 05765-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RT-SB-01-10-11

Quant Time: Dec 18 00:43:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121523S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 16 00:39:52 2023
Response via : Initial Calibration

