

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121623\
 Data File : VY016674.D
 Acq On : 16 Dec 2023 17:13
 Operator : SY/MD
 Sample : 05758-01RE
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SP-COMP-1

Quant Time: Dec 18 00:42:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121523S.M
 Quant Title : SW846 8260
 QLast Update : Sat Dec 16 00:39:52 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.795	168	103915	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	215971	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	279593	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	113980	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	63406	74.487	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 148.980%		
35) Dibromofluoromethane	7.728	113	66405	57.322	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 114.640%		
50) Toluene-d8	10.185	98	336367	72.665	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	= 145.340%#		
62) 4-Bromofluorobenzene	12.483	95	106037	71.425	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	= 142.860%		
Target Compounds						
						Qvalue
20) Methylene Chloride	4.729	84	16046	10.957	ug/l #	82


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121623\
Data File : VY016674.D
Acq On : 16 Dec 2023 17:13
Operator : SY/MD
Sample : 05758-01RE
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SP-COMP-1

Quant Time: Dec 18 00:42:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121523S.M
Quant Title : SW846 8260
QLast Update : Sat Dec 16 00:39:52 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.795 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY016674.D
Acq: 16 Dec 2023 17:13




Instrument :
MSVOA_Y
ClientSampleId :
SP-COMP-1


Tgt Ion:168 Resp: 103915
Ion Ratio Lower Upper
168 100
99 49.8 43.4 65.0



#20
Methylene Chloride
Concen: 10.957 ug/l
RT: 4.729 min Scan# 477
Delta R.T. -0.006 min
Lab File: VY016674.D
Acq: 16 Dec 2023 17:13

Tgt Ion: 84 Resp: 16046
Ion Ratio Lower Upper
84 100
49 102.5 106.9 160.3#
51 32.4 33.2 49.8#
86 65.4 51.0 76.6





#33

1,2-Dichloroethane-d4

Concen: 74.487 ug/l

RT: 8.149 min Scan# 110

Delta R.T. -0.000 min

Lab File: VY016674.D

Acq: 16 Dec 2023 17:13

Instrument :

MSVOA_Y

ClientSampleId :





SP-COMP-1

Tgt Ion: 65 Resp: 63406

Ion Ratio Lower Upper

65 100

67 54.7 0.0 101.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.697 min Scan# 1128

Delta R.T. -0.006 min

Lab File: VY016674.D

Acq: 16 Dec 2023 17:13




Tgt Ion: 114 Resp: 215971


Ion Ratio Lower Upper

114 100

63 17.8 0.0 42.4

88 12.7 0.0 30.6









#35
 Dibromofluoromethane
 Concen: 57.322 ug/l
 RT: 7.728 min Scan# 90
 Delta R.T. 0.000 min
 Lab File: VY016674.D
 Acq: 16 Dec 2023 17:13

Instrument :
 MSVOA_Y
 ClientSampleId :
 SP-COMP-1

Tgt Ion:113 Resp: 66405




Ion	Ratio	Lower	Upper
113	100		
111	102.1	81.3	121.9
192	19.4	16.9	25.3




#50
 Toluene-d8
 Concen: 72.665 ug/l
 RT: 10.185 min Scan# 1372
 Delta R.T. -0.006 min
 Lab File: VY016674.D
 Acq: 16 Dec 2023 17:13

Tgt Ion: 98 Resp: 336367

Ion	Ratio	Lower	Upper
98	100		
100	64.6	50.1	75.1









#62
4-Bromofluorobenzene
Concen: 71.425 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. -0.006 min
Lab File: VY016674.D
Acq: 16 Dec 2023 17:13




Instrument :
MSVOA_Y
ClientSampleId :
SP-COMP-1


Tgt Ion: 95 Resp: 106037
Ion Ratio Lower Upper
95 100
174 93.6 0.0 178.2
176 89.7 0.0 173.8



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.496 min Scan# 1587
Delta R.T. 0.000 min
Lab File: VY016674.D
Acq: 16 Dec 2023 17:13

Tgt Ion: 117 Resp: 279593
Ion Ratio Lower Upper
117 100
82 51.4 43.4 65.0
119 33.0 26.3 39.5





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.428 min Scan# 1904

Delta R.T. 0.000 min

Lab File: VY016674.D


Acq: 16 Dec 2023 17:13

Instrument :

MSVOA_Y

ClientSampleId :

SP-COMP-1



Tgt Ion:152 Resp: 113980

Ion Ratio Lower Upper

152 100

115 56.3 28.8 86.5

150 157.9 0.0 348.8

