

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121625\
 Data File : VY023991.D
 Acq On : 16 Dec 2025 15:39
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Dec 17 04:14:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121625S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 17 04:12:33 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	692039	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	1145187	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	992449	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	406845	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	329550	63.007	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	126.020%
35) Dibromofluoromethane	7.640	113	363955	53.106	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.220%
50) Toluene-d8	10.109	98	1295787	52.399	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	104.800%
62) 4-Bromofluorobenzene	12.407	95	391054	48.247	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	96.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121625\
Data File : VY023991.D
Acq On : 16 Dec 2025 15:39
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Dec 17 04:14:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121625S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 17 04:12:33 2025
Response via : Initial Calibration

