

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY121820\
 Data File : VY003817.D
 Acq On : 18 Dec 2020 14:39
 Operator : SY/MD
 Sample : L5151-04
 Misc : 4.44G/10ML/MSVOA Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 C0AE6

Quant Time: Dec 19 00:52:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\SOM2YLM121420S.M
 Quant Title : VOC Analysis
 QLast Update : Sat Dec 19 00:17:11 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.70	114	198584	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.49	117	165237	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.43	152	57419	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.25	65	67016	24.73	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	98.92%
7) Chloroethane-d5	2.77	69	53720	26.59	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	106.36%
10) 1,1-Dichloroethene-d2	3.86	63	92191	16.19	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	64.76%
20) 2-Butanone-d5	6.91	46	35707	30.23	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	60.46%
24) Chloroform-d	7.49	84	121767	23.61	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	94.44%
26) 1,2-Dichloroethane-d4	8.15	65	64205	22.02	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	88.08%
29) Benzene-d6	8.12	84	268082	28.03	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	112.12%
33) 1,2-Dichloropropane-d6	9.12	67	79099	26.56	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	106.24%
37) Toluene-d8	10.18	98	236568	27.33	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	109.32%
38) trans-1,3-Dichloropropene-	10.44	79	32617	22.47	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	89.88%
39) 2-Hexanone-d5	10.79	63	28154	34.28	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	68.56%
48) 1,1,2,2-Tetrachloroethane-	12.56	84	52834	20.29	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	81.16%
61) 1,2-Dichlorobenzene-d4	13.72	152	55489	26.90	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	107.60%

Target Compounds					Ovalue
16) Methylene chloride	4.74	84	12128	3.402 ug/L	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY121820\
Data File : VY003817.D
Acq On : 18 Dec 2020 14:39
Operator : SY/MD
Sample : L5151-04
Misc : 4.44G/10ML/MSVOA Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
C0AE6

Quant Time: Dec 19 00:52:00 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\SOM2YLM121420S.M
Quant Title : VOC Analysis
QLast Update : Sat Dec 19 00:17:11 2020
Response via : Initial Calibration

