

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_Y\DATA\VY121919\
 Data File : VY001012.D
 Acq On : 19 Dec 2019 13:58
 Operator : SY/MD
 Sample : K6370-07RE
 Misc : 3.57G/5ML/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Quant Time: Dec 20 04:12:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y112719S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 27 12:40:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.80	168	222506	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.70	114	374942	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.49	117	233655	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.42	152	60054	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.15	65	107745	45.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.14%	
35) Dibromofluoromethane	7.73	113	112300	51.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.78%	
50) Toluene-d8	10.18	98	419881	48.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.12%	
62) 4-Bromofluorobenzene	12.48	95	73666	22.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	44.90%	
Target Compounds						
30) Chloroform	7.52	83	13907	2.925	ug/l	98
32) 1,1,1-Trichloroethane	7.71	97	26399	6.725	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_Y\DATA\VY121919\
Data File : VY001012.D
Acq On : 19 Dec 2019 13:58
Operator : SY/MD
Sample : K6370-07RE
Misc : 3.57G/5ML/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Quant Time: Dec 20 04:12:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\82Y112719S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 27 12:40:47 2019
Response via : Initial Calibration

