

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121923\
 Data File : VY016695.D
 Acq On : 19 Dec 2023 13:12
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Dec 20 05:31:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 20 05:26:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	107140	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	186403	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	177883	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	72114	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	58648	66.180	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	132.360%
35) Dibromofluoromethane	7.728	113	62837	57.237	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.480%
50) Toluene-d8	10.191	98	226223	56.898	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	113.800%
62) 4-Bromofluorobenzene	12.489	95	62233	55.050	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	110.100%
Target Compounds						Qvalue
20) Methylene Chloride	4.728	84	20483	17.891	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121923\
Data File : VY016695.D
Acq On : 19 Dec 2023 13:12
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Dec 20 05:31:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121923S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 20 05:26:40 2023
Response via : Initial Calibration

