

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121923\
 Data File : VY016700.D
 Acq On : 19 Dec 2023 18:04
 Operator : SY/MD
 Sample : VY1219SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1219SBL01

Quant Time: Dec 20 06:36:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 20 05:26:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	172732	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.703	114	264813	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	217043	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	91262	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	62405	43.679	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	87.360%
35) Dibromofluoromethane	7.734	113	67864	43.512	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	87.020%
50) Toluene-d8	10.191	98	249842	44.232	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	88.460%
62) 4-Bromofluorobenzene	12.489	95	61833	38.501	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	77.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121923\
Data File : VY016700.D
Acq On : 19 Dec 2023 18:04
Operator : SY/MD
Sample : VY1219SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1219SBL01

Quant Time: Dec 20 06:36:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121923S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 20 05:26:40 2023
Response via : Initial Calibration

