

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121923\
 Data File : VY016703.D
 Acq On : 19 Dec 2023 19:16
 Operator : SY/MD
 Sample : 05796-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RT-SB-04-10-11

Quant Time: Dec 20 06:38:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 20 05:26:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	98172	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.710	114	177086	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.502	117	171565	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	71198	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	56887	70.056	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 140.120%	
35) Dibromofluoromethane	7.734	113	59343	56.898	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 113.800%	
50) Toluene-d8	10.191	98	214531	56.796	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 113.600%	
62) 4-Bromofluorobenzene	12.489	95	59435	55.341	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 110.680%	
Target Compounds						
17) Carbon Disulfide	4.210	76	2315	1.058	ug/l	# 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121923\
Data File : VY016703.D
Acq On : 19 Dec 2023 19:16
Operator : SY/MD
Sample : 05796-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RT-SB-04-10-11

Quant Time: Dec 20 06:38:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121923S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 20 05:26:40 2023
Response via : Initial Calibration

