

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121923\
 Data File : VY016704.D
 Acq On : 19 Dec 2023 19:40
 Operator : SY/MD
 Sample : 05796-02
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RT-SB-05-10.25-11.25

Quant Time: Dec 20 06:38:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 20 05:26:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	93782	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.704	114	168654	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	160851	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	67523	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	52113	67.181	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 134.360%	
35) Dibromofluoromethane	7.734	113	32842	33.063	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 66.120%	
50) Toluene-d8	10.191	98	204743	56.915	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 113.820%	
62) 4-Bromofluorobenzene	12.489	95	58222	56.922	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 113.840%	
Target Compounds						Qvalue
17) Carbon Disulfide	4.216	76	6734	3.221	ug/l #	94
52) Toluene	10.252	92	3468	1.313	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121923\
Data File : VY016704.D
Acq On : 19 Dec 2023 19:40
Operator : SY/MD
Sample : 05796-02
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RT-SB-05-10.25-11.25

Quant Time: Dec 20 06:38:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121923S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 20 05:26:40 2023
Response via : Initial Calibration

