

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121923\
 Data File : VY016706.D
 Acq On : 19 Dec 2023 20:26
 Operator : SY/MD
 Sample : 05796-04
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RT-SB-04-13-14

Quant Time: Dec 20 06:39:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 20 05:26:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	71321	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.703	114	120019	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	103241	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	35310	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	37009	62.735	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	125.480%
35) Dibromofluoromethane	7.734	113	40455	57.231	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.460%
50) Toluene-d8	10.191	98	139571	54.520	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	109.040%
62) 4-Bromofluorobenzene	12.495	95	34625	47.569	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	95.140%
Target Compounds						
17) Carbon Disulfide	4.216	76	6927	4.357	ug/l	Qvalue # 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121923\
Data File : VY016706.D
Acq On : 19 Dec 2023 20:26
Operator : SY/MD
Sample : 05796-04
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RT-SB-04-13-14

Quant Time: Dec 20 06:39:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121923S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 20 05:26:40 2023
Response via : Initial Calibration

