

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121923\
 Data File : VY016712.D
 Acq On : 19 Dec 2023 22:47
 Operator : SY/MD
 Sample : 05827-03
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RT-SB-07-10-11

Quant Time: Dec 20 06:41:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 20 05:26:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	69778	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	123977	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.507	117	116595	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.440	152	45062	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.161	65	41610	72.094	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 144.180%	
35) Dibromofluoromethane	7.734	113	43699	59.847	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 119.700%	
50) Toluene-d8	10.197	98	148986	56.340	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	= 112.680%	
62) 4-Bromofluorobenzene	12.495	95	38712	51.486	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 102.980%	
Target Compounds						Qvalue
20) Methylene Chloride	4.734	84	9552	11.182	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121923\
Data File : VY016712.D
Acq On : 19 Dec 2023 22:47
Operator : SY/MD
Sample : 05827-03
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RT-SB-07-10-11

Quant Time: Dec 20 06:41:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121923S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 20 05:26:40 2023
Response via : Initial Calibration

