

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121923\
 Data File : VY016717.D
 Acq On : 20 Dec 2023 00:44
 Operator : SY/MD
 Sample : 05827-08
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RT-SB-08-15.5-16.5

Quant Time: Dec 20 06:43:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 20 05:26:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.820	168	23085	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.716	114	37750	50.000	ug/l	0.02
63) Chlorobenzene-d5	11.514	117	34133	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.447	152	12440	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.167	65	14250	74.629	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	= 149.260%	
35) Dibromofluoromethane	7.740	113	14369	64.628	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 129.260%	
50) Toluene-d8	10.203	98	44939	55.811	ug/l	0.02
Spiked Amount	50.000	Range	58 - 134	Recovery	= 111.620%	
62) 4-Bromofluorobenzene	12.502	95	10670	46.605	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	= 93.220%	
Target Compounds						
20) Methylene Chloride	4.747	84	2353	6.860	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121923\
Data File : VY016717.D
Acq On : 20 Dec 2023 00:44
Operator : SY/MD
Sample : 05827-08
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RT-SB-08-15.5-16.5

Quant Time: Dec 20 06:43:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121923S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 20 05:26:40 2023
Response via : Initial Calibration

