

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121924\
 Data File : VY020646.D
 Acq On : 19 Dec 2024 11:07
 Operator : SY/MD
 Sample : VY1219SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1219SBL01

Quant Time: Dec 20 00:46:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121724S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 18 05:27:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	204045	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	318952	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	271132	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	145317	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	99418	44.488	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	88.980%
35) Dibromofluoromethane	7.634	113	102958	45.802	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.600%
50) Toluene-d8	10.109	98	344522	43.788	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	87.580%
62) 4-Bromofluorobenzene	12.408	95	124147	40.337	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	80.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121924\
Data File : VY020646.D
Acq On : 19 Dec 2024 11:07
Operator : SY/MD
Sample : VY1219SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1219SBL01

Quant Time: Dec 20 00:46:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121724S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 18 05:27:13 2024
Response via : Initial Calibration

