

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121924\
 Data File : VY020655.D
 Acq On : 19 Dec 2024 15:12
 Operator : SY/MD
 Sample : P5306-07
 Misc : 5.31g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-VSL-10-121224

Quant Time: Dec 20 00:55:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121724S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 18 05:27:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	12480	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	15453	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	10205	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	1903	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	6253	45.748	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.500%
35) Dibromofluoromethane	7.634	113	6171	56.662	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.320%
50) Toluene-d8	10.109	98	19631	51.498	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.000%
62) 4-Bromofluorobenzene	12.408	95	3853	25.839	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	51.680%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121924\
Data File : VY020655.D
Acq On : 19 Dec 2024 15:12
Operator : SY/MD
Sample : P5306-07
Misc : 5.31g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-VSL-10-121224

Quant Time: Dec 20 00:55:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121724S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 18 05:27:13 2024
Response via : Initial Calibration

