



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121924\
Data File : VY020644.D
Acq On : 19 Dec 2024 09:13
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121724S.M
Title : SW846 8260
Last Update : Wed Dec 18 05:27:13 2024



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1744

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.1	27309	PASS
75	95	30	60	50.7	76547	PASS
95	95	100	100	100.0	150976	PASS
96	95	5	9	6.5	9845	PASS
173	174	0.00	2	1.1	1249	PASS
174	95	50	100	78.7	118747	PASS
175	174	5	9	7.6	8988	PASS
176	174	95	101	96.2	114192	PASS
177	176	5	9	6.9	7838	PASS