

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122024\
 Data File : VY020674.D
 Acq On : 20 Dec 2024 16:31
 Operator : SY/MD
 Sample : P5343-01
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ210

Quant Time: Dec 20 18:26:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121724S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 18 05:27:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	845	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.616	114	1237	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	393	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.359	152	22	50.000	ug/l	# 0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	476	51.434	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.860%
35) Dibromofluoromethane	7.634	113	478	54.829	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.660%
50) Toluene-d8	10.109	98	953	31.231	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	62.460%
62) 4-Bromofluorobenzene	12.408	95	22	1.843	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	3.680%#
Target Compounds						
16) Acetone	3.860	43	23	6.234	ug/l #	43
76) 1,2,3-Trichloropropane	12.408	75	65	284.417	ug/l #	100
81) trans-1,4-Dichloro-2-b...	12.408	75	65	597.004	ug/l #	12

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122024\
Data File : VY020674.D
Acq On : 20 Dec 2024 16:31
Operator : SY/MD
Sample : P5343-01
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ210

Quant Time: Dec 20 18:26:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121724S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 18 05:27:13 2024
Response via : Initial Calibration

