

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122123\
 Data File : VY016773.D
 Acq On : 21 Dec 2023 21:41
 Operator : SY/MD
 Sample : 05878-03
 Misc : 4.15g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COLZ3

Quant Time: Dec 22 00:46:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 21 01:49:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.813	168	113730	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.716	114	234677	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.508	117	211565	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.446	152	74038	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.167	65	58298	56.151	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.300%
35) Dibromofluoromethane	7.746	113	60446	44.965	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.940%
50) Toluene-d8	10.203	98	212949	43.943	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	87.880%
62) 4-Bromofluorobenzene	12.501	95	67381	42.858	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	85.720%
Target Compounds						
16) Acetone	3.973	43	37156	132.345	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122123\
Data File : VY016773.D
Acq On : 21 Dec 2023 21:41
Operator : SY/MD
Sample : 05878-03
Misc : 4.15g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COLZ3

Quant Time: Dec 22 00:46:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 21 01:49:45 2023
Response via : Initial Calibration

