

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122123\
 Data File : VY016777.D
 Acq On : 22 Dec 2023 00:20
 Operator : SY/MD
 Sample : VY1221SBL02
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1221SBL02

Quant Time: Dec 22 00:47:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 21 01:49:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.819	168	128818	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.716	114	245965	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.514	117	213498	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.447	152	84482	50.000	ug/l	0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.173	65	76179	64.780	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	129.560%
35) Dibromofluoromethane	7.746	113	77234	54.817	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.640%
50) Toluene-d8	10.203	98	263194	51.819	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.640%
62) 4-Bromofluorobenzene	12.502	95	80241	48.695	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	97.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122123\
Data File : VY016777.D
Acq On : 22 Dec 2023 00:20
Operator : SY/MD
Sample : VY1221SBL02
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1221SBL02

Quant Time: Dec 22 00:47:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 21 01:49:45 2023
Response via : Initial Calibration

