

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122123\
 Data File : VY016781.D
 Acq On : 22 Dec 2023 01:51
 Operator : SY/MD
 Sample : 05876-09
 Misc : 5.93g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COLX3

Quant Time: Dec 22 03:18:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 21 01:49:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.819	168	115866	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.722	114	234815	50.000	ug/l	0.02
63) Chlorobenzene-d5	11.514	117	214457	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.446	152	74613	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.173	65	57328	54.199	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.400%
35) Dibromofluoromethane	7.752	113	62538	46.494	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.980%
50) Toluene-d8	10.203	98	227961	47.013	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.020%
62) 4-Bromofluorobenzene	12.508	95	68706	43.675	ug/l	0.02
Spiked Amount	50.000	Range	30 - 143	Recovery	=	87.360%
Target Compounds						
20) Methylene Chloride	4.759	84	13264	6.341	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122123\
Data File : VY016781.D
Acq On : 22 Dec 2023 01:51
Operator : SY/MD
Sample : 05876-09
Misc : 5.93g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COLX3

Quant Time: Dec 22 03:18:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 21 01:49:45 2023
Response via : Initial Calibration

