

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122123\
 Data File : VY016786.D
 Acq On : 22 Dec 2023 03:48
 Operator : SY/MD
 Sample : 05876-14
 Misc : 2.65g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COLX9

Quant Time: Dec 22 06:41:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 21 01:49:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.826	168	113885	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.722	114	230620	50.000	ug/l	0.02
63) Chlorobenzene-d5	11.514	117	200980	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.447	152	66530	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.173	65	55561	53.442	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.880%
35) Dibromofluoromethane	7.752	113	59863	45.315	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.620%
50) Toluene-d8	10.209	98	213780	44.891	ug/l	0.02
Spiked Amount	50.000	Range	58 - 134	Recovery	=	89.780%
62) 4-Bromofluorobenzene	12.508	95	63860	41.333	ug/l	0.02
Spiked Amount	50.000	Range	30 - 143	Recovery	=	82.660%
Target Compounds						
11) Tert butyl alcohol	4.991	59	3331	39.077	ug/l	Qvalue # 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122123\
Data File : VY016786.D
Acq On : 22 Dec 2023 03:48
Operator : SY/MD
Sample : 05876-14
Misc : 2.65g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COLX9

Quant Time: Dec 22 06:41:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 21 01:49:45 2023
Response via : Initial Calibration

