

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122123\
 Data File : VY016788.D
 Acq On : 22 Dec 2023 04:35
 Operator : SY/MD
 Sample : 05876-16
 Misc : 5.44g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COLY1

Quant Time: Dec 22 06:41:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 21 01:49:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.819	168	92628	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.716	114	193938	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.514	117	181523	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.447	152	64460	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.173	65	55646	65.807	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	= 131.620%	
35) Dibromofluoromethane	7.746	113	59236	53.321	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.640%	
50) Toluene-d8	10.203	98	214078	53.456	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	= 106.920%	
62) 4-Bromofluorobenzene	12.508	95	66371	51.083	ug/l	0.02
Spiked Amount	50.000	Range	30 - 143	Recovery	= 102.160%	
Target Compounds						
20) Methylene Chloride	4.753	84	10019	5.828	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122123\
Data File : VY016788.D
Acq On : 22 Dec 2023 04:35
Operator : SY/MD
Sample : 05876-16
Misc : 5.44g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COLY1

Quant Time: Dec 22 06:41:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 21 01:49:45 2023
Response via : Initial Calibration

