

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122123\
 Data File : VY016790.D
 Acq On : 22 Dec 2023 05:22
 Operator : SY/MD
 Sample : 05876-18
 Misc : 4.02g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COLY3

Quant Time: Dec 22 06:41:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 21 01:49:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.825	168	120895	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.722	114	247266	50.000	ug/l	0.02
63) Chlorobenzene-d5	11.514	117	227480	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.446	152	83791	50.000	ug/l	0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.173	65	54619	49.490	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.980%
35) Dibromofluoromethane	7.752	113	60461	42.686	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	85.380%
50) Toluene-d8	10.209	98	215160	42.139	ug/l	0.02
Spiked Amount	50.000	Range	58 - 134	Recovery	=	84.280%
62) 4-Bromofluorobenzene	12.502	95	66104	39.905	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	79.820%

Target Compounds						Qvalue
30) Chloroform	7.539	83	3715	1.528	ug/l	# 78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122123\
Data File : VY016790.D
Acq On : 22 Dec 2023 05:22
Operator : SY/MD
Sample : 05876-18
Misc : 4.02g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COLY3

Quant Time: Dec 22 06:41:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 21 01:49:45 2023
Response via : Initial Calibration

