

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122123\
 Data File : VY016799.D
 Acq On : 22 Dec 2023 08:53
 Operator : SY/MD
 Sample : 05877-25
 Misc : 2.45g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COM66

Quant Time: Dec 22 09:02:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 21 01:49:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.825	168	121907	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.722	114	248472	50.000	ug/l	0.02
63) Chlorobenzene-d5	11.514	117	225189	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.446	152	80112	50.000	ug/l	0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.173	65	60004	53.918	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.840%
35) Dibromofluoromethane	7.758	113	61494	43.205	ug/l	0.03
Spiked Amount	50.000	Range	54 - 147	Recovery	=	86.400%
50) Toluene-d8	10.209	98	216637	42.222	ug/l	0.02
Spiked Amount	50.000	Range	58 - 134	Recovery	=	84.440%
62) 4-Bromofluorobenzene	12.508	95	67064	40.288	ug/l	0.02
Spiked Amount	50.000	Range	30 - 143	Recovery	=	80.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122123\
Data File : VY016799.D
Acq On : 22 Dec 2023 08:53
Operator : SY/MD
Sample : 05877-25
Misc : 2.45g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COM66

Quant Time: Dec 22 09:02:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 21 01:49:45 2023
Response via : Initial Calibration

