

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122123\
 Data File : VY016797.D
 Acq On : 22 Dec 2023 08:06
 Operator : SY/MD
 Sample : 05877-23
 Misc : 4.36g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COM62

Quant Time: Dec 22 08:52:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 21 01:49:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.826	168	87868	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.722	114	170849	50.000	ug/l	0.02
63) Chlorobenzene-d5	11.514	117	147984	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.453	152	47072	50.000	ug/l	0.02

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.173	65	41605	51.867	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.740%
35) Dibromofluoromethane	7.752	113	41956	42.871	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	85.740%
50) Toluene-d8	10.209	98	144829	41.051	ug/l	0.02
Spiked Amount	50.000	Range	58 - 134	Recovery	=	82.100%
62) 4-Bromofluorobenzene	12.508	95	41969	36.667	ug/l	0.02
Spiked Amount	50.000	Range	30 - 143	Recovery	=	73.340%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122123\
Data File : VY016797.D
Acq On : 22 Dec 2023 08:06
Operator : SY/MD
Sample : 05877-23
Misc : 4.36g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COM62

Quant Time: Dec 22 08:52:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 21 01:49:45 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.826 min Scan# 91
Delta R.T. 0.025 min
Lab File: VY016797.D
Acq: 22 Dec 2023 08:06




Instrument :
MSVOA_Y
ClientSampleId :
COM62


Tgt Ion:168 Resp: 87868
Ion Ratio Lower Upper
168 100
99 57.0 44.2 66.4



#33
1,2-Dichloroethane-d4
Concen: 51.867 ug/l
RT: 8.173 min Scan# 1042
Delta R.T. 0.019 min
Lab File: VY016797.D
Acq: 22 Dec 2023 08:06

Tgt Ion: 65 Resp: 41605
Ion Ratio Lower Upper
65 100
67 54.5 0.0 109.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.722 min Scan# 1129

Delta R.T. 0.019 min

Lab File: VY016797.D

Acq: 22 Dec 2023 08:06

Instrument :

MSVOA_Y

ClientSampleId :

COM62





Tgt Ion:114 Resp: 170849

Ion Ratio Lower Upper

114 100

63 19.8 0.0 37.4

88 14.5 0.0 29.0



#35

Dibromofluoromethane

Concen: 42.871 ug/l

RT: 7.752 min Scan# 973

Delta R.T. 0.025 min

Lab File: VY016797.D

Acq: 22 Dec 2023 08:06




Tgt Ion:113 Resp: 41956


Ion Ratio Lower Upper

113 100

111 103.4 83.8 125.6

192 17.9 14.5 21.7









#50
Toluene-d8
Concen: 41.051 ug/l
RT: 10.209 min Scan# 11
Delta R.T. 0.018 min
Lab File: VY016797.D
Acq: 22 Dec 2023 08:06




Instrument :
MSVOA_Y
ClientSampleId :
COM62


Tgt Ion: 98 Resp: 144829
Ion Ratio Lower Upper
98 100
100 65.5 52.3 78.5



#62
4-Bromofluorobenzene
Concen: 36.667 ug/l
RT: 12.508 min Scan# 1753
Delta R.T. 0.018 min
Lab File: VY016797.D
Acq: 22 Dec 2023 08:06

Tgt Ion: 95 Resp: 41969
Ion Ratio Lower Upper
95 100
174 79.0 0.0 160.0
176 75.1 0.0 151.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.514 min Scan# 11

Delta R.T. 0.012 min

Lab File: VY016797.D

Acq: 22 Dec 2023 08:06

Instrument :

MSVOA_Y

ClientSampleId :

COM62





Tgt Ion:117 Resp: 147984

Ion Ratio Lower Upper

117 100

82 56.2 43.8 65.8

119 32.5 26.5 39.7



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.453 min Scan# 1908

Delta R.T. 0.018 min

Lab File: VY016797.D

Acq: 22 Dec 2023 08:06

Tgt Ion:152 Resp: 47072

Ion Ratio Lower Upper

152 100

115 59.9 30.0 90.0

150 163.1 0.0 344.6

