

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122123\
 Data File : VY016799.D
 Acq On : 22 Dec 2023 08:53
 Operator : SY/MD
 Sample : 05877-25
 Misc : 2.45g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COM66

Quant Time: Dec 22 09:02:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 21 01:49:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.825	168	121907	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.722	114	248472	50.000	ug/l	0.02
63) Chlorobenzene-d5	11.514	117	225189	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.446	152	80112	50.000	ug/l	0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.173	65	60004	53.918	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.840%
35) Dibromofluoromethane	7.758	113	61494	43.205	ug/l	0.03
Spiked Amount	50.000	Range	54 - 147	Recovery	=	86.400%
50) Toluene-d8	10.209	98	216637	42.222	ug/l	0.02
Spiked Amount	50.000	Range	58 - 134	Recovery	=	84.440%
62) 4-Bromofluorobenzene	12.508	95	67064	40.288	ug/l	0.02
Spiked Amount	50.000	Range	30 - 143	Recovery	=	80.580%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122123\
Data File : VY016799.D
Acq On : 22 Dec 2023 08:53
Operator : SY/MD
Sample : 05877-25
Misc : 2.45g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COM66

Quant Time: Dec 22 09:02:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 21 01:49:45 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.825 min Scan# 91
Delta R.T. 0.024 min
Lab File: VY016799.D
Acq: 22 Dec 2023 08:53




Instrument :
MSVOA_Y
ClientSampleId :
COM66


Tgt Ion:168 Resp: 121907
Ion Ratio Lower Upper
168 100
99 57.5 44.2 66.4



#33
1,2-Dichloroethane-d4
Concen: 53.918 ug/l
RT: 8.173 min Scan# 1042
Delta R.T. 0.018 min
Lab File: VY016799.D
Acq: 22 Dec 2023 08:53

Tgt Ion: 65 Resp: 60004
Ion Ratio Lower Upper
65 100
67 55.2 0.0 109.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.722 min Scan# 1129

Delta R.T. 0.019 min

Lab File: VY016799.D

Acq: 22 Dec 2023 08:53

Instrument :

MSVOA_Y

ClientSampleId :

COM66





Tgt Ion:114 Resp: 248472

Ion Ratio Lower Upper

114 100

63 19.8 0.0 37.4

88 14.5 0.0 29.0



#35

Dibromofluoromethane

Concen: 43.205 ug/l

RT: 7.758 min Scan# 974

Delta R.T. 0.031 min

Lab File: VY016799.D

Acq: 22 Dec 2023 08:53




Tgt Ion:113 Resp: 61494


Ion Ratio Lower Upper

113 100

111 101.3 83.8 125.6

192 17.6 14.5 21.7









#50
Toluene-d8
Concen: 42.222 ug/l
RT: 10.209 min Scan# 11
Delta R.T. 0.018 min
Lab File: VY016799.D
Acq: 22 Dec 2023 08:53




Instrument :
MSVOA_Y
ClientSampleId :
COM66


Tgt Ion: 98 Resp: 216637
Ion Ratio Lower Upper
98 100
100 64.9 52.3 78.5



#62
4-Bromofluorobenzene
Concen: 40.288 ug/l
RT: 12.508 min Scan# 1753
Delta R.T. 0.018 min
Lab File: VY016799.D
Acq: 22 Dec 2023 08:53

Tgt Ion: 95 Resp: 67064
Ion Ratio Lower Upper
95 100
174 80.6 0.0 160.0
176 76.6 0.0 151.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.514 min Scan# 11

Delta R.T. 0.012 min

Lab File: VY016799.D

Acq: 22 Dec 2023 08:53

Instrument :

MSVOA_Y

ClientSampleId :

COM66





Tgt Ion:117 Resp: 225189

Ion Ratio Lower Upper

117 100

82 56.5 43.8 65.8

119 31.9 26.5 39.7



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.446 min Scan# 1907

Delta R.T. 0.012 min

Lab File: VY016799.D

Acq: 22 Dec 2023 08:53

Tgt Ion:152 Resp: 80112

Ion Ratio Lower Upper

152 100

115 60.4 30.0 90.0

150 156.8 0.0 344.6

