

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122223\
 Data File : VY016804.D
 Acq On : 22 Dec 2023 13:24
 Operator : SY/MD
 Sample : VY1222SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1222SBL01

Quant Time: Dec 22 22:57:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 21 01:49:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.813	168	106722	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.716	114	232236	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.514	117	236832	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.446	152	88187	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.167	65	71716	73.611	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	147.220%
35) Dibromofluoromethane	7.740	113	74736	56.180	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.360%
50) Toluene-d8	10.203	98	273669	57.067	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	114.140%
62) 4-Bromofluorobenzene	12.502	95	89289	57.390	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	114.780%
Target Compounds						
20) Methylene Chloride	4.747	84	14543	8.116	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122223\
Data File : VY016804.D
Acq On : 22 Dec 2023 13:24
Operator : SY/MD
Sample : VY1222SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1222SBL01

Quant Time: Dec 22 22:57:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 21 01:49:45 2023
Response via : Initial Calibration

