

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122223\
 Data File : VY016832.D
 Acq On : 23 Dec 2023 01:40
 Operator : SY/MD
 Sample : VY1222SBL02
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1222SBL02

Quant Time: Dec 23 01:56:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 21 01:49:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.825	168	120062	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.722	114	233684	50.000	ug/l	0.02
63) Chlorobenzene-d5	11.514	117	203676	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.446	152	76031	50.000	ug/l	0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.173	65	71204	64.965	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	129.920%
35) Dibromofluoromethane	7.752	113	70837	52.919	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.840%
50) Toluene-d8	10.209	98	240358	49.810	ug/l	0.02
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.620%
62) 4-Bromofluorobenzene	12.507	95	72932	46.586	ug/l	0.02
Spiked Amount	50.000	Range	30 - 143	Recovery	=	93.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122223\
Data File : VY016832.D
Acq On : 23 Dec 2023 01:40
Operator : SY/MD
Sample : VY1222SBL02
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1222SBL02

Quant Time: Dec 23 01:56:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 21 01:49:45 2023
Response via : Initial Calibration

