

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122223\
 Data File : VY016836.D
 Acq On : 23 Dec 2023 03:57
 Operator : SY/MD
 Sample : 05957-08
 Misc : 4.56g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COM35

Quant Time: Dec 23 05:35:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 21 01:49:45 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.825	168	110178	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.722	114	240675	50.000	ug/l	0.02
63) Chlorobenzene-d5	11.514	117	241142	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.446	152	90219	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.173	65	70619	70.211	ug/l	0.02
Spiked Amount	50.000	Range 50 - 163	Recovery	= 140.420%		
35) Dibromofluoromethane	7.752	113	73699	53.458	ug/l	0.02
Spiked Amount	50.000	Range 54 - 147	Recovery	= 106.920%		
50) Toluene-d8	10.203	98	267207	53.765	ug/l	0.01
Spiked Amount	50.000	Range 58 - 134	Recovery	= 107.540%		
62) 4-Bromofluorobenzene	12.508	95	86322	53.537	ug/l	0.02
Spiked Amount	50.000	Range 30 - 143	Recovery	= 107.080%		
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.759	84	11622	5.610	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122223\
Data File : VY016836.D
Acq On : 23 Dec 2023 03:57
Operator : SY/MD
Sample : 05957-08
Misc : 4.56g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COM35

Quant Time: Dec 23 05:35:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 21 01:49:45 2023
Response via : Initial Calibration

