

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122223\
 Data File : VY016838.D
 Acq On : 23 Dec 2023 04:44
 Operator : SY/MD
 Sample : 05957-10
 Misc : 3.02g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COM46

Quant Time: Dec 23 05:36:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 21 01:49:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.826	168	104772	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.722	114	233558	50.000	ug/l	0.02
63) Chlorobenzene-d5	11.514	117	232152	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.447	152	87698	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.173	65	64741	67.688	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	135.380%
35) Dibromofluoromethane	7.752	113	70468	52.671	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.340%
50) Toluene-d8	10.209	98	258735	53.647	ug/l	0.02
Spiked Amount	50.000	Range	58 - 134	Recovery	=	107.300%
62) 4-Bromofluorobenzene	12.508	95	82527	52.743	ug/l	0.02
Spiked Amount	50.000	Range	30 - 143	Recovery	=	105.480%
Target Compounds						
20) Methylene Chloride	4.765	84	10361	5.076	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122223\
Data File : VY016838.D
Acq On : 23 Dec 2023 04:44
Operator : SY/MD
Sample : 05957-10
Misc : 3.02g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COM46

Quant Time: Dec 23 05:36:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 21 01:49:45 2023
Response via : Initial Calibration

