

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122223\
 Data File : VY016840.D
 Acq On : 23 Dec 2023 05:31
 Operator : SY/MD
 Sample : 05957-12
 Misc : 3.10g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COM60

Quant Time: Dec 23 06:09:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 21 01:49:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.819	168	52500	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.722	114	108359	50.000	ug/l	0.02
63) Chlorobenzene-d5	11.514	117	105901	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.446	152	37578	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.173	65	31991	66.750	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	133.500%
35) Dibromofluoromethane	7.752	113	34302	55.263	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.520%
50) Toluene-d8	10.209	98	119085	53.220	ug/l	0.02
Spiked Amount	50.000	Range	58 - 134	Recovery	=	106.440%
62) 4-Bromofluorobenzene	12.508	95	35201	48.490	ug/l	0.02
Spiked Amount	50.000	Range	30 - 143	Recovery	=	96.980%
Target Compounds						
16) Acetone	3.979	43	2550	19.676	ug/l	90
20) Methylene Chloride	4.759	84	4486	3.989	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122223\
Data File : VY016840.D
Acq On : 23 Dec 2023 05:31
Operator : SY/MD
Sample : 05957-12
Misc : 3.10g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COM60

Quant Time: Dec 23 06:09:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122023S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 21 01:49:45 2023
Response via : Initial Calibration

